

ENGLISH

For use with HM-JACK series

Wash Liquid (Auto Detergent H)

IVD CE

For Professional Use Only

[PACKAGE CONTENTS]

Code	Product Name	Constituent**	Quantity
REF 052404	Wash Liquid (Auto Detergent H)	Potassium hydroxide (<5%), Nitrilotriacetic acid trisodium salt (5-15%), Antiseptic (<0.3%), Water and other constituents (79-89%)	200 mL x 1

[INTENDED USE]

This product is washing solution used for cleaning the nozzles of fully automated immunoassay analyzer**, HM-JACK series.

[PROCEDURE]

- Pipette 1.0 mL of the Wash Liquid (Auto Detergent H) into designated Wash Container (Wash Tank).
- Dilute with 1.0 L of distilled or de-ionized water to prepare 1:1000 solution (Wash Solution). Keep the same dilution ratio in order to prepare large amount of Wash Solution if necessary.
- Mix this Wash Solution gently but thoroughly and leave for 20 minutes at room temperature.
- Set the prepared Wash Solution on the HM-JACK series analyzer and run WASH** prior to measurement. Refer to the operation manual for more detail.
- Use prepared Wash Solution within 30 days and DO NOT add freshly prepared Wash Solution to remaining.

[MATERIAL REQUIRED BUT NOT PROVIDED]

- Pipette
- Wash Tank

[GENERAL WARNINGS AND PRECAUTIONS]

- Do not attempt to pipette by mouth.
- Do not ingest.
- Do not inhale. Operate in a well-ventilated area.
- Avoid contact with skin and eyes. Wear protective clothes and equipment when handling.
- Avoid contact with acidic substances to prevent heat generation.**
- Avoid contact with non-ferrous metals (such as, copper, aluminium and brass) other than stainless steel to prevent corrosion.**
- Process the waste appropriately according to local biological waste regulations.
- Keep the product under the indicated conditions before the expiry.
- Do not use the product that has expired its shelf life.**
- Prepare the Wash Solutions following the given instructions. Alterations of the specified procedure may affect the optimal performance of the product.
- Do not use the product for any other purpose.
- Do not use if there is damage or leakage observed.**
- This product is specially designed for use with fully automated immunoassay analyzer, HM-JACK series. Do not use with any other analyzers.**
- This product contains potassium hydroxide and Nitrilotriacetic acid trisodium salt.
 - Skin corrosion/irritation : Category 1
 - Serious eye damage /eye irritation : Category 1
 - Carcinogenicity : Category 2



[HAZARD STATEMENT]

H314 Causes severe skin burns and eye damage.
H318 Causes serious eye damage.
H351 Suspected of causing cancer

[PRECAUTIONARY STATEMENT]

Prevention

P201 Obtain special instructions before use.**
P202 Do not handle until all safety precautions have been read and understood.
P260 Do not breathe dust/fume/gas/mist/vapors/spray.
P264 Wash contaminated parts thoroughly after handling.
P280 Wear protective gloves, protective clothing or eye protection/face protection.

Response

P308 + P313 IF exposed or concerned: Get medical advice/attention.

P310 Immediately call a POISON CENTER or doctor/physician.

P304 + P340 IF INHALED: Remove victim to fresh air and keep at rest in a position comfortable for breathing.

P303 + P361 + P353 IF ON SKIN (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower.

P363 Wash contaminated clothing before reuse.

P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P301 + P330 + P331 IF SWALLOWED: Rinse mouth. Do NOT induce vomiting.

Storage

P405 Store locked up.

Disposal

P501 Dispose of contents/container in accordance with local/national regulation.














[STORAGE AND SHELF LIFE]

1. Storage: Do not freeze. Store at 2-35°C
2. Shelf life: 60 months

[STORAGE AFTER ONCE OPENING]**

1. After once opening, be sure to store at 2-35°C with the lid closed and use as soon as possible.

[DEFINITION OF SYMBOLS]**

Symbol	Definition	Symbol	Definition
	<i>In vitro</i> diagnostic medical device		Temperature limitation
	Manufacturer		Use by
	CE mark		Authorized Representative in the European Community
	Consult instructions for use		Skin corrosion
	Contents		Carcinogenicity
	Catalog number		For professional use only
	Batch/Lot code		

For all inquiries, contact:**

Minaris Medical Co., Ltd.

Quality Assurance and Regulatory Affairs Department

1-8-10, Harumi, Chuo-ku, Tokyo, 104-6004 JAPAN

Please report any serious incident associated with the product to the above address and the competent authority of the Member State.**



Minaris Medical Co., Ltd.*

1-8-10, Harumi, Chuo-ku,

Tokyo, 104-6004 JAPAN

Tel: +81-3-6219-7617 — Fax: +81-3-6219-7614



Obelis s.a.**

Bd. Général Wahis 53

B-1030 Brussels, Belgium